

ATTESTATION OF CONFORMITY

Certificate Number	: TD10502301
Manufacturer's Name	: Topair Systems, Inc.
Manufacturer's Address	: 300 First Avenue, Suite 102 Needham MA 02494 USA
Product Name	: Biological Safety Cabinet
Model-Type	: BO-090-PP,BO-090-ME,BO-090-PP-B,BO-090-ME-B,BO-120-PP,BO-120-ME, BO-120-PP-B,BO-120-ME-B,BO-150-PP,BO-150-ME,BO-150-PP-B,BO-150-ME-B, BO-180-PP,BO-180-ME,BO-180-PP-B,BO-180-ME-B,ECO-BO-070-PP,ECO-BO-080-PP, BO-090-NSF,BO-120-NSF,BO-150-NSF,BO-180-NSF,BO-090-EN-IPMS,BO-120-EN-IPMS, BO-150-EN-IPMS,BO-180-EN-IPMS,BO-090-EN,BO-120-EN,BO-150-EN,BO-180-EN, BO-090-EN-B2,BO-120-EN-B2,BO-150-EN-B2,BO-180-EN-B2
Reference Directive	: 2014/35/EU Low Voltage Directive 2014/30/EU Electromagnetic Compatibility
Test Report No.	: TS-23000420-1 TS-23000420-2
Reference Standard(s)	: EN 61010-1:2010+A1:2019 EN IEC 61326-1:2021 EN IEC 61000-3-2:2019 EN 61000-3-3:2013+A2:2021

Result:

The technical file defined in Annex-II of 2014/30/EU Electromagnetic Compatibility & Annex-III of 2014/35/EU Low Voltage Directive has been prepared by the manufacturer for the above-mentioned product. Compliance of the product and documentation with the EU directives and standards is the responsibility of the manufacturer. Any significant changes in design and/or production of the product or amendments to the relevant EU directives or standards referred above may render this certificate invalid. The validity of this certificate can be verified on www.eutest.com.tr

Certificate issue date : 24.04.2023

Certificate expiry date : 23.04.2028




Hikmet Elp
General Manager